

## SPQ PROTEIN

---

### CARDIAC RISK ASSESSMENT

#### APO A-1 SPQ II Kit 86070 Liquid

|                               |                                                                                      |
|-------------------------------|--------------------------------------------------------------------------------------|
| <b>Intended Use</b>           | Quantitative determination of APO A-1 by immunoprecipitin methods.                   |
| <b>Method</b>                 | Nephelometric/Turbidimetric                                                          |
| <b>No. Standard</b>           | 5                                                                                    |
| <b>Kit Components</b>         | Antibody Reagent Set II for APO A-1 (86059). APO A-1 Calibration/Control Set (86102) |
| <b>Feature &amp; Benefits</b> | Liquid reagents. Additional control material may be purchased separately.            |
| <b>Sample Types</b>           | Serum/EDTA-Plasma                                                                    |

#### APO A-1 Antibody Reagent Set II 86059 Liquid

|                               |                                                                                                                                                       |
|-------------------------------|-------------------------------------------------------------------------------------------------------------------------------------------------------|
| <b>Intended Use</b>           | The SPQ Antibody Reagent Set II for Apolipoprotein A-1 is designed for the quantitative determination of APO A-1 levels by immunoprecipitin analysis. |
| <b>Method</b>                 | Nephelometric/Turbidimetric                                                                                                                           |
| <b>Kit Components</b>         | APO A-1 Antibody Reagent. APO A-1 Antibody Diluent. Diluent A                                                                                         |
| <b>Feature &amp; Benefits</b> | Liquid reagents. Additional control material may be purchased separately.                                                                             |
| <b>Sample Types</b>           | Serum/EDTA-Plasma                                                                                                                                     |

#### APO A-1 Calibration/Control Set 86102 Liquid

|                               |                                                                                                                                                                                             |
|-------------------------------|---------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
| <b>Intended Use</b>           | The SPQ Calibration/Control Set for Apolipoprotein A-1 (APO A-1) is designed for the quantitative determination of APO A-1 in conjunction with the SPQ Antibody Reagent Set II for APO A-1. |
| <b>Method</b>                 | Nephelometric/Turbidimetric                                                                                                                                                                 |
| <b>No. Standard</b>           | 5                                                                                                                                                                                           |
| <b>Kit Components</b>         | Calibration Set for Apolipoprotein A-1. Control for APO A-1                                                                                                                                 |
| <b>Feature &amp; Benefits</b> | Liquid reagents. Additional control material may be purchased separately.                                                                                                                   |
| <b>Sample Types</b>           | Serum                                                                                                                                                                                       |

### APO B SPQ II Kit 86071 Liquid

|                               |                                                                                 |
|-------------------------------|---------------------------------------------------------------------------------|
| <b>Intended Use</b>           | Quantitative determination of APO B by immunoprecipitin methods.                |
| <b>Method</b>                 | Nephelometric/Turbidimetric                                                     |
| <b>No. Standard</b>           | 5                                                                               |
| <b>Kit Components</b>         | Antibody Reagent Set II for APO B (86060). APO B Calibrator/Control Set (86103) |
| <b>Feature &amp; Benefits</b> | Liquid reagents. Components may also be ordered separately.                     |
| <b>Sample Types</b>           | Serum/EDTA-Plasma                                                               |

### APO B Antibody Reagent Set II 86060 Liquid

|                               |                                                                                                                                                   |
|-------------------------------|---------------------------------------------------------------------------------------------------------------------------------------------------|
| <b>Intended Use</b>           | The SPQ Antibody Reagent Set II for Apolipoprotein B is designed for the quantitative determination of APO B levels by immunoprecipitin analysis. |
| <b>Method</b>                 | Nephelometric/Turbidimetric                                                                                                                       |
| <b>Kit Components</b>         | APO B Antibody Reagent. Polymer Diluent. Diluent A                                                                                                |
| <b>Feature &amp; Benefits</b> | Liquid reagents.                                                                                                                                  |
| <b>Sample Types</b>           | Serum/EDTA-Plasma                                                                                                                                 |

### APO B Calibration/Control Set 86103 Liquid

|                               |                                                                                                                                                                                     |
|-------------------------------|-------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
| <b>Intended Use</b>           | The SPQ Calibration/Control Set for Apolipoprotein B (APO B) is designed for the quantitative determination of APO B in conjunction with the SPQ Antibody Reagent Set II for APO B. |
| <b>Method</b>                 | Nephelometric/Turbidimetric                                                                                                                                                         |
| <b>No. Standard</b>           | 5                                                                                                                                                                                   |
| <b>Kit Components</b>         | Calibration Set for Apo-B. Control for Apo-B                                                                                                                                        |
| <b>Feature &amp; Benefits</b> | Liquid reagents. Additional control material may be purchased separately.                                                                                                           |
| <b>Sample Types</b>           | Serum                                                                                                                                                                               |

### Fibrinogen SPQ II Kit 86113 Liquid

|                               |                                                                                            |
|-------------------------------|--------------------------------------------------------------------------------------------|
| <b>Intended Use</b>           | Quantitative determination of Fibrinogen by immunoprecipitin methods.                      |
| <b>Method</b>                 | Nephelometric/Turbidimetric                                                                |
| <b>No. Standard</b>           | 5 (serial dilution)                                                                        |
| <b>Kit Components</b>         | Antibody Reagent Set II for Fibrinogen (86086). Fibrinogen Calibration/Control Set (86116) |
| <b>Feature &amp; Benefits</b> | Components may also be ordered separately.                                                 |
| <b>Sample Types</b>           | Citrated Plasma/EDTA-Plasma                                                                |

**Fibrinogen Antibody Reagent Set II**

86086 Liquid

|                               |                                                                                                                                                     |
|-------------------------------|-----------------------------------------------------------------------------------------------------------------------------------------------------|
| <b>Intended Use</b>           | The SPQ Antibody Reagent Set II for Fibrinogen is designed for the quantitative determination of fibrinogen levels by immunoprecipitation analysis. |
| <b>Method</b>                 | Nephelometric/Turbidimetric                                                                                                                         |
| <b>Kit Components</b>         | Fibrinogen Antibody Reagent. Polymer Diluent III. Diluent A                                                                                         |
| <b>Feature &amp; Benefits</b> | Liquid reagents. Test systems may also be ordered as the separate kit components.                                                                   |
| <b>Sample Types</b>           | Citrated Plasma/EDTA-Plasma                                                                                                                         |

**Fibrinogen Calibration/Control Set**

86116 Lyophilized

|                               |                                                                                                                                                                                          |
|-------------------------------|------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
| <b>Intended Use</b>           | The SPQ Calibration/Control Set for Fibrinogen is designed to assist in the quantitative determination of fibrinogen in conjunction with the SPQ Antibody Reagent Set II for Fibrinogen. |
| <b>Method</b>                 | Nephelometric/Turbidimetric                                                                                                                                                              |
| <b>No. Standard</b>           | 5 (serial dilution)                                                                                                                                                                      |
| <b>Kit Components</b>         | Calibrator for Fibrinogen. Control for Fibrinogen                                                                                                                                        |
| <b>Feature &amp; Benefits</b> | Liquid reagents. Additional control material may be purchased separately.                                                                                                                |
| <b>Sample Types</b>           | Citrated Plasma/EDTA-Plasma                                                                                                                                                              |

**Lipoprotein(a) SPQ II Kit**

86111 Liquid

|                               |                                                                                  |
|-------------------------------|----------------------------------------------------------------------------------|
| <b>Intended Use</b>           | Quantitative determination of Lp(a) by immunoprecipitin methods.                 |
| <b>Method</b>                 | Nephelometric/Turbidimetric                                                      |
| <b>No. Standard</b>           | 5                                                                                |
| <b>Kit Components</b>         | Antibody Reagent Set II for Lp(a) (86084). Lp(a) Calibration/Control Set (86114) |
| <b>Feature &amp; Benefits</b> | Liquid reagents. Components may also be ordered separately.                      |
| <b>Sample Types</b>           | Serum/EDTA-Plasma                                                                |

**Lipoprotein(a) Antibody Reagent Set II**

86084 Liquid

|                               |                                                                                                                                        |
|-------------------------------|----------------------------------------------------------------------------------------------------------------------------------------|
| <b>Intended Use</b>           | The SPQ Antibody Reagent Set II for Lp(a) is designed for the quantitative determination of Lp(a) levels by immunoprecipitin analysis. |
| <b>Method</b>                 | Nephelometric/Turbidimetric                                                                                                            |
| <b>Kit Components</b>         | Lp(a) Antibody Reagent. Lp(a) Diluent                                                                                                  |
| <b>Feature &amp; Benefits</b> | Liquid reagents. Test systems may also be ordered as the separate kit components. No sample pre-treatment is necessary.                |
| <b>Sample Types</b>           | Serum/EDTA-Plasma                                                                                                                      |

**Lipoprotein(a) Calibration/Control Set**  
86114 Liquid

|                               |                                                                                                                                                                                                |
|-------------------------------|------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
| <b>Intended Use</b>           | The SPQ Calibration/Control Set for Lp(a) is designed for quantitative determination of Lipoprotein(a) in human serum or plasma in conjunction with the SPQ Antibody Reagent Set II for Lp(a). |
| <b>Method</b>                 | Nephelometric/Turbidimetric                                                                                                                                                                    |
| <b>No. Standard</b>           | 5                                                                                                                                                                                              |
| <b>Kit Components</b>         | Calibration Set for Lipoprotein(a). Controls for Lp(a)                                                                                                                                         |
| <b>Feature &amp; Benefits</b> | Liquid reagents. No sample pre-treatment required. Additional control material may be purchased separately.                                                                                    |
| <b>Sample Types</b>           | Serum/EDTA-Plasma                                                                                                                                                                              |