

## INFECTIOUS DISEASE

---

### OTHER INFECTIOUS DISEASES - ELISA

#### ETI - Varicella Zoster (VZV) IgG 4621/96 Wells

<b>Intended Use</b>	For the Detection of IgG Antibodies to Varicella Zoster Virus
<b>Label</b>	HRP
<b>Method</b>	ELISA
<b>Controls Included</b>	One Calibrator and Two Controls (Positive and Negative)
<b>Incubations</b>	30 min + 30 min + 30 min/18-30°C
<b>Chromogene/Substrate</b>	TMB
<b>Sample Types</b>	Serum
<b>Sample Size</b>	5 µL of (1:101) dilution
<b>Comments</b>	In Vitro Diagnostic Use

#### ETI-Measles IgG 4601/96 Wells

<b>Intended Use</b>	For the Detection of IgG antibodies to Measles Virus
<b>Label</b>	HRP
<b>Method</b>	ELISA
<b>Controls Included</b>	One calibrator and two controls (Positive and Negative)
<b>Incubations</b>	30 min + 30 min + 30 min/18-30°C
<b>Chromogene/Substrate</b>	TMB
<b>Sample Types</b>	Serum
<b>Sample Size</b>	5 µL of (1:101) dilution
<b>Comments</b>	In Vitro Diagnostic Use

#### H. Pylori IgG 4651/96 Wells

<b>Intended Use</b>	For the Detection of IgG Antibodies to H. Pylori
<b>Label</b>	HRP
<b>Method</b>	ELISA
<b>Controls Included</b>	One cut-off calibrator and two controls (Positive and Negative)
<b>Incubations</b>	30 min + 30 min/RT + 30 min/18-30°C
<b>Chromogene/Substrate</b>	TMB
<b>Sample Types</b>	Serum
<b>Sample Size</b>	10 µL (1:21 dilution)
<b>Comments</b>	In Vitro Diagnostic Use

### Mumps IgG Elisa 4631/96 Wells

<b>Intended Use</b>	For the Detection of IgG antibodies to Mumps Virus
<b>Label</b>	HRP
<b>Method</b>	ELISA
<b>Controls Included</b>	One cut-off calibrator and two controls (Positive and Negative)
<b>Incubations</b>	30 mins + 30 mins + 30 min/18-30°C
<b>Chromogene/Substrate</b>	TMB
<b>Sample Types</b>	Serum
<b>Sample Size</b>	10 µL of 1:21 dilution
<b>Comments</b>	In Vitro Diagnostic Use

### TREP-SURE™ Syphilis Screen TS960/960 Wells

<b>Intended Use</b>	For the in vitro diagnostic detection of Treponema Pallidum Antibodies
<b>Label</b>	HRP
<b>Method</b>	ELISA
<b>Controls Included</b>	One cut-off calibrator and two controls (Positive, Negative)
<b>Incubations</b>	60 min + 30 min + 15 min/37°C
<b>Chromogene/Substrate</b>	TMB
<b>Sample Types</b>	Serum or EDTA and Citrated Plasma
<b>Sample Size</b>	100 µL
<b>Comments</b>	In Vitro Diagnostic Use.

### TREP-SURE™ Syphilis Screen TS96/96 Wells

<b>Intended Use</b>	For the in vitro diagnostic detection of Treponema Pallidum Antibodies.
<b>Label</b>	HRP
<b>Method</b>	ELISA
<b>Controls Included</b>	One cut-off calibrator and two controls (Positive, Negative)
<b>Incubations</b>	60 min + 30 min + 15 min/37°C
<b>Chromogene/Substrate</b>	TMB
<b>Sample Types</b>	Serum or EDTA and Citrated Plasma
<b>Sample Size</b>	100 µL
<b>Comments</b>	In Vitro Diagnostic Use.